

# **EZ1030 Phenol Analyser**

Method and reagent sheets 04/2023, Edition 2.01

## **Table of Contents**

1.	Legal information	3
2.	Analytical specifications	3
3.	Analysis method	4
4.	Reagents	5
4.1	Reagent overview and consumption	6
	DI-water overview and consumption	
	Storage and quality of chemicals	
	Buffer solution	
	Colour solution	
	Calibration solution	
	Cleaning solution (facultative)	
r.,	Ciculing Jointon (radaltative)	د

## 1. Legal information

Manufacturer: AppliTek NV/SA

Distributor: Hach Lange GmbH

The translation of the manual is approved by the manufacturer.

## 2. Analytical specifications

Please refer also to the respective technical datasheet at Hach Support Online.

Phenol - All specifications	S					
Analysis method	Colorimetric measurement at 505 nm using 4-aminoantipyrine					
Parameter Phenol						
	Stand	lard measurement cycle time: 10 minutes				
Cycle time		al dilution: + 5 min.				
		nal dilution: + 5 – 10 min.				
Limit of detection (LOD)	≤ 0.0′	I mg/L				
Precision/Repeatability	Bette	than 2% full scale range for standard test s	olutions			
Cleaning	Auton	natic; frequency freely programmable				
Calibration	Auton	natic, 2-point; frequency freely programmabl	е			
Validation	Auton	natic; frequency freely programmable				
Interferences		Oxidizing agents such as chlorine (Cl <sub>2</sub> ); sulphur compounds like sulphide [(S) <sup>2-</sup> ]. Large amounts of color and turbidity interfere. Fats, oil, proteins, surfactants and tar.				
Measuring ranges	% of	range - Dilution	Low range (mg/L)	High range (mg/L)		
			\ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \ \	\		
	А	10% of standard range	0.01	0.2		
	A B	10% of standard range 25% of standard range				
		,	0.01	0.2		
	В	25% of standard range	0.01 0.025	0.2		
	В	25% of standard range 50% of standard range	0.01 0.025 0.025	0.2 0.5 1		
	B C <b>0</b>	25% of standard range 50% of standard range standard range	0.01 0.025 0.025 <b>0.1</b>	0.2 0.5 1 2		
	B C <b>0</b>	25% of standard range 50% of standard range standard range internal MP dilution (factor 4)	0.01 0.025 0.025 <b>0.1</b> 0.8	0.2 0.5 1 2 8		
	B C 0 1	25% of standard range 50% of standard range standard range internal MP dilution (factor 4) Internal MP dilution (factor 8)	0.01 0.025 0.025 <b>0.1</b> 0.8 1.6	0.2 0.5 1 2 8 16		
	B C 0 1 2 W	25% of standard range 50% of standard range standard range internal MP dilution (factor 4) Internal MP dilution (factor 8) internal dispenser dilution (factor 10)	0.01 0.025 0.025 0.1 0.8 1.6 1	0.2 0.5 1 2 8 16 20		
	B C 0 1 2 W X	25% of standard range 50% of standard range standard range internal MP dilution (factor 4) Internal MP dilution (factor 8) internal dispenser dilution (factor 10) internal dispenser dilution (factor 25)	0.01 0.025 0.025 0.1 0.8 1.6 1 2,5	0.2 0.5 1 2 8 16 20 50		

### 3. Analysis method

#### **Summary**

The determination of the phenol concentration in water is based on the reaction of phenol with 4-aminoantipyrine in the presence of potassium ferrocyanide to form an intense coloured reddish-brown complex. The absorption is measured at 505 nm.

#### **Analysis steps**

The analysis vessel is cleaned and filled with fresh sample. After sampling, the initial absorbance value is measured at 505 nm. Next, buffer solution and colour solution are added and after respecting a stirring period – performed to obtain complete colour development –the final absorbance value is determined. With the obtained absorbance values, the phenol concentration can be calculated according to Beer's Law.

#### Calibration

The calibration procedure measures a REF1 Phenol solution (channel 9, REF1 valve) and a REF2 Phenol solution (channel 10, REF2 valve) to adapt the slope and offset factors by means of a two point calibration.

The calibration is performed in the MAIN method.

#### Remark

The methods cannot be started at the same time.

## 4. Reagents

## **A** CAUTION



Chemical exposure hazard. Obey laboratory safety procedures and wear all of the personal protective equipment appropriate to the chemicals that are handled. Read the safety data sheet from the supplier before bottles are filled or reagents are prepared. For laboratory use only. Make the hazard information known in accordance with the local regulations of the user.

## **A** CAUTION



Chemical exposure hazard. Dispose of chemicals and wastes in accordance with local, regional and national regulations.

## 4.1 Reagent overview and consumption

In the tables below, the products that are needed to prepare the reagents are listed. The product name, the formula, the molecular weight, the CAS No. and the amount needed to prepare 1 liter of the reagents is given. Check the consumption of the reagents (28 days) to adapt the volumes needed.

Product	Consumption	Consumption/28 days A rata 1 analysis/10 min	Recommended containers
Buffer solution	~ 0.5 mL / analysis	~ 2.0 L	Plastic – 2.5 L
Colour solution	~ 0.5 mL / analysis	~ 2.0 L	Glass – 2.5 L
REF1 solution	~ 0.5 L / calibration	1	Plastic – 1 L
REF2 solution	~ 0.5 L / calibration	1	Plastic – 1 L

## 4.2 DI-water overview and consumption

	Rinse water (mL/analysis) Type I	Dilution water (mL/analysis) Type I	Total (mL/analysis)	Consumption/28 days A rata 1 analysis / 10 min
Α	N.A.	N.A.	N.A.	N.A.
В	N.A.	N.A.	N.A.	N.A.
С	N.A.	N.A.	N.A.	N.A.
0	N.A.	N.A.	N.A.	N.A.
1	60 mL	15 mL	75 mL	302 L
2	60 mL	15 mL	75 mL	302 L
W	60 mL	15 mL	75 mL	302 L
Х	60 mL	15 mL	75 mL	302 L
Υ	60 mL	15 mL	75 mL	302 L
Z	60 mL	15 mL	75 mL	302 L
5	60 mL	15 mL	75 mL	302 L

#### Remark

The indicated volumes are an estimation of the consumption for rinse and dilution water, based on a standard operating procedure, as defined in the specifications of the EZ analyser. Please be aware that, depending on the sample matrix, the rinse water volumes might increase.

### 4.3 Storage and quality of chemicals

#### **Quality of chemicals**

All chemicals should be of Reagent grade, ACS grade or better (\*). The use of pro analysis chemicals is recommended. Poor quality of the reagents can affect the analyser performance.

(\*) Analytical Reagent (AR), Guaranteed Reagent (GR), UNIVAR, AnalaR, Premium Reagent (PR), ReagentCertified ACS reagent, ACS Plus reagent, puriss p.a. ACS reagent, ReagentPlus®, TraceCERT®, Suprapur®, Ultrapur®, or better are also possible.

#### **Quality of DI-water**

All EZ analysers are tested with standard solutions, reagents and dilution water prepared using type I water or better as defined by ASTM D1193-91.

To achieve the specifications as stated on the data sheet, method and reagents sheet and acceptance test reports, the same water quality (or better) must be used for the preparation of the standard solutions, reagents and dilution water.

Additionally the water used for the preparation of the standard solutions for an EZ analyser must be free of the parameter or any of the interferences for the method of that EZ analyser.

#### Storage of Reagents

While operating the instrument, keep in mind the reagent requirements as stated in the reagent overview, the chapters below and/or in the data sheet of the instrument.

### **A CAUTION**



For longer-term storage: Store the reagents cold; Store the reagents in the dark;

If applicable: Store the reagents in a fridge during operation

## **A** CAUTION



Refresh the reagents after one month (unless stated differently in the chapters below).

Do not mix old reagents with freshly prepared reagents. Remove old reagents from the container before adding freshly prepared reagents.

### 4.4 Buffer solution

Products	Formula	MW (g/mol)	CAS No.	1 litre solution
Potassium hexacyanoferrate (III)	K <sub>3</sub> Fe(CN) <sub>6</sub>	329.24	13746-66-2	15.385 g
Ammonium hydroxide solution (25%)*	NH₄OH	35.05	1336-21-6	77 mL

<sup>\*</sup> Density: 0.91 g/ml (20°C)

#### Preparation

Dissolve 15.385 g of potassium hexacyanoferrate ( $K_3Fe(CN)_6$ ) in 300 mL de-ionized water. Add 77 mL ammonium hydroxide solution (NH<sub>4</sub>OH, 25%). Mix and fill up to 1 litre with de-ionized water.

### 4.5 Colour solution

Products	Formula	MW (g/mol)	CAS No.	1 litre solution
4-amino-antipyrine	C <sub>11</sub> H <sub>13</sub> N <sub>3</sub> O	203.24	83-07-8	16.67 g

#### Preparation

Dissolve 16.67 g of 4-amino-antipyrine ( $C_{11}H_{13}N_3O$ ) in 500 mL de-ionized water. Mix and fill up to 1 litre de-ionized water

#### 4.6 Calibration solution

Products	Formula	MW (g/mol)	CAS No.	1 litre solution
Phenol	C <sub>6</sub> H <sub>6</sub> O	94.11	108-95-2	1 g

#### **Preparation**

#### 1000 mg/L Phenol stock solution

Prepare a stock solution of 1000 mg/L Phenol: Dissolve accurately 1 g phenol ( $C_6H_6O$ ) in 500 mL de-ionized water using a volumetric flask of 1000 mL. Fill up to 1 litre with de-ionized water.

#### Phenol standard solution - REF2

Prepare a standard solution for calibration according to the following table: take accurately x mL of the 1000 mg/L Phenol stock solution and transfer into a volumetric flask of 1000 mL. Add de-ionized water up to the mark grade.

	Measuring range	Concentration REF2	Amount of stock solution to add to 1 litre
Α	0.2 mg/L Phenol	0.2 mg/L Phenol	0.2 mL
В	0.5 mg/L Phenol	0.5 mg/L Phenol	0.5 mL
С	1.0 mg/L Phenol	1.0 mg/L Phenol	1.0 mL
0	2.0 mg/L Phenol	2.0 mg/L Phenol	2.0 mL
1	8.0 mg/L Phenol	8.0 mg/L Phenol	8 mL
2	16.0 mg/L Phenol	16.0 mg/L Phenol	16 mL
W	20 mg/L Phenol	20 mg/L Phenol	20 mL
Х	50 mg/L Phenol	50 mg/L Phenol	50 mL
Υ	100 mg/L Phenol	100 mg/L Phenol	100 mL
Z	150 mg/L Phenol	150 mg/L Phenol	150 mL
5	200 mg/L Phenol	200 mg/L Phenol	200 mL

#### Phenol standard solution - REF1

Prepare a standard solution of 0 mg/L Phenol. Use de-ionized water.

## 4.7 Cleaning solution (facultative)

The cleaning procedure should prevent any build-up of chemicals in the analyser. To obtain an effective cleaning procedure one has to test the cleaning solution and the cleaning interval for each application. Perform the selected cleaning solution and interval for a trial period, check then the effectiveness of the procedure and change if necessary.

	Change Information					
Date: 26/04/2023	Previous version: Edition 1.01 to Edition 2.01					
Reason for Change						
	ake in documentation change the range of A ake in documentation change the LOD of 100X dilution					

### **Description of Change**

- Change the range from 0.25 ppm to 0.2 ppm of the 10% of the standard range Change LOD of the 100X dilution from 2 to 10 mg/L  $\,$