




### Test preparation


**CAUTION:** *Review the Safety Data Sheets (MSDS/SDS) for the chemicals that are used. Use the recommended personal protective equipment.*

- Put the color disc on the center pin in the color comparator box (numbers to the front).
- Use the indoor light color disc when the light source is fluorescent light. Use the outdoor light color disc when the light source is sunlight.
- Rinse the tubes with sample before the test. Rinse the tubes with deionized water after the test.
- If the color match is between two segments, use the value that is in the middle of the two segments.
- If the color disc becomes wet internally, pull apart the flat plastic sides to open the color disc. Remove the thin inner disc. Dry all parts with a soft cloth. Assemble when fully dry.
- Undissolved reagent does not have an effect on test accuracy.
- To verify the test accuracy, use a standard solution as the sample.
- If the sample contains rust or precipitated iron, fully mix the sample and then fill the tubes. Wait 2–5 minutes after the reagent is added. Dissolved iron develops a color immediately.
- Samples that contain more than 4 mg/L iron can give low results. If high iron levels are possible, dilute the sample as follows. Use a 3-mL syringe to add 2.5 mL of sample to each tube. Dilute the sample to the 5-mL mark with deionized water. Use the diluted sample in the test procedure and multiply the result by 2. Use the syringe to add 1 mL of sample to each tube. Dilute the sample to the 5-mL mark with deionized water. Use the diluted sample in the test procedure and multiply the result by 5.


### Test procedure—Iron (0–4 mg/L Fe)




1. Fill two tubes to the first line (5 mL) with sample.




2. Put one tube into the left opening of the color comparator box.




3. Add one FerroVer Iron Reagent Powder Pillow to the second tube.




4. Swirl to mix. An orange color develops.



5. Put the second tube into the color comparator box.



6. Hold the color comparator box in front of a light source. Turn the color disc to find the color match.



7. Read the result in mg/L in the scale window.

### Replacement items

Description	Unit	Item no.
FerroVer® Iron Reagent Powder Pillows, 5 mL	100/pkg	92799
Color disc, iron, indoor light, 0–4 mg/L	each	9262400
Color disc, iron, outdoor light, 0–4 mg/L	each	9263800
Color comparator box	each	173200
Plastic viewing tubes, 18 mm, with caps	4/pkg	4660004

### Optional items

Description	Unit	Item no.
Caps for plastic viewing tubes (4660004)	4/pkg	4660014
Water, deionized	500 mL	27249
Glass viewing tubes, glass, 18 mm	6/pkg	173006
Iron standard solution, 1 mg/L Fe	500 mL	13949
Stoppers for 18-mm glass tubes and AccuVac Ampuls	6/pkg	173106
Syringe, Luer-Lok® Tip, 3 mL	each	4321300

