



# Quanta™

Performance that repeats



Volume   
Viscosity 

Device Type	Automatic, multi-dose pen
Primary Container	1.5 mL and 3.0 mL RTU cartridges according to ISO 13926-1
Dose Type	Multi-fixed-dose   4-6 doses
Dose volume	Up to 0.75 mL
Route of Administration	Subcutaneous
Activation Method	Spring-assisted button
Injection Feedback	Audible click at start and near end of injection, viewing window, dose indicator

Copyright © 2026 SHL Medical AG. All rights reserved. Version 1.0 Information accurate at time of print and is subject to change without notice. Contact [info@shl-medical.com](mailto:info@shl-medical.com) for more information.

## Handling Instructions



1  
Uncap device and attach needle onto pen.



2  
Rotate dialing knob until activation button appears.



3  
Press the button and hold in position to inject.



+  
Remove used needle and discard.

## Key Features

- ✓ No manual priming required
- ✓ Activation button only appears once the set dose is dialed
- ✓ Built-in indicator confirms when the device is ready to inject
- ✓ No dose alterations possible by patient