





Device Type Disposable autoinjector

Primary Container 5.0 mL cartridge with pre-installed needle (NIT®)

Route of Administration Subcutaneous

Dosage Type Fixed dose | Optional automated priming and titration

Volume Viscosity \_\_\_

Activation Method Needle cover

Injection Feedback Continuous audible clicks during injection, viewing window

\* For suspension drugs that require a priming step - hold device with cap upright during cap removal.

Copyright © 2025 SHL Medical AG. All rights reserved. Version 1.1 Information accurate at time of print and is subject to change without notice. Contact info@shl-medical.com for more information.

## **Handling Instructions**



Unscrew cap until needle cover automatically extends\*. Then remove cap.



Press and hold against injection site to activate.



Discard in sharps container.

## **Key Features**

- √ Simple 2-step operation
- ✓ Needle cover activation
- ✓ Continuous audible and visual feedback
- ✓ One-handed operation
- ✓ Permanently hidden needle

- ✓ Compact design
- ✓ Customizable needle length and gauge
- √ Passive locking needle cover
- ✓ Customized to client specifications

