

CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

Hardy Diagnostics 1430 W. McCoy Lane Santa Maria, CA 93455

Fulfills the requirements of

ISO/IEC 17025:2017

In the field of

CALIBRATION

This certificate is valid only when accompanied by a current scope of accreditation document. The current scope of accreditation can be verified at www.anab.org.

Jason Stine, Vice President

Expiry Date: 16 January 2028 Certificate Number: AC-2546





SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

Hardy Diagnostics

1430 W. McCoy Lane Santa Maria, CA 93455

Rachel Ma 805-346-27<mark>66X5</mark>679 MaR@hardydiagnostics.com www.hardydiagnostics.com

CALIBRATION

ISO/IEC 17025 Accreditation Granted: 24 December 2025

Certificate Number: AC-2546 Certificate Expiry Date: 16 January 2028

Mass and Mass Related

Parameter/Equipment	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Air Samplers (Velocity) ¹	25 LPM	2 LPM	Wind Tunnel, Reference
	100 LPM	4 LPM	Anemometer, Barometer,
	200 LPM	5 LPM	Hygrometer

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 (*k*=2), corresponding to a confidence level of approximately 95%.

Notes:

1. On-site calibration service is available for this parameter, since on-site conditions are typically more variable than those in the laboratory, larger measurement uncertainties are expected on-site than what is reported on the accredited scope





